

ICSI Dish
(#20253)

1. Description

ICSI Dish is designed for precise manipulation of reproductive cells, such as in Intracytoplasmic Sperm Injection(ICSI) procedures.

2. Features

- ❖ Low-wall and External grip for better handling
- ❖ USP guideline compliant (USP class VI tested)
- ❖ Mouse Embryo Assay (MEA) test passed
- ❖ Human Sperm Survival Assay (HSSA) test passed: SMI ≥ 0.75

3. Functional Test

Sterility	:	Passed
Non-cytotoxic	:	Passed
Non-pyrogenic	:	Passed
DNase/RNase free	:	Passed
Human DNA free	:	Passed

4. Product Range

Cat.No.	Material	External Dimension (mm)	Surface Treatment	Sterile	Packaging
20253	PS	50.00 x 9.00	-	+	10 / 120

5. Certification


SPL Life Sciences hereby certifies that the product identified above, inspected to be in compliance with product quality specification and requirements as documented in our ISO 9001:2015 Quality Management System (K-QA-Q031478) in Korea.

6. Specifications & Technical Drawing

- Specifications

Dimensions				Material (Resin)	Packaging
Lid		Dish			
Diameter Outer:	51.72 mm	Diameter Grip:	50.70 mm	Crystal-grade polystyrene (PS)	▪ Pieces/Sleeve: 10 ▪ Pieces/Case: 120
Diameter Inner:	49.85 mm	Diameter Dish Outer:	49.00 mm		
Height:	3.80 mm	Diameter Dish Inner:	47.60 mm		
		Height:	8.85 mm		

- Technical Drawing



www.spllifesciences.com

For technical assistance, contact SPL R&D Center at:
Tel: +82-31-533-4800; Fax: +82-31-533-1430; e-mail: spl@ispl.co.kr
To place an order, contact your local distributor or
Tel: +82-31-533-4800; Fax: +82-31-533-1430; e-mail: business@ispl.co.kr